LK7349-LAS

0046376

Lockheed Environmental Systems & Technologies Co.
Lockheed Analytical Services
975 Kelly Johnson Drive Las Vegas, Nevada 89119-3705
Telephone 702-361-0220 800-582-7605 Facsimile 702-361-8146

LOCKHEED MARTIN

July 25, 1996

Ms. Joan Kessner
Bechtel Hanford, Inc.
3350 George Washington Way
MISN B1-35
Richland, WA 99352

RE:	Log-in No.:	L7349
	Quotation No.:	Q400000-B
	SAF:	B96-092
	Document File No.:	0629596
	BHI Document File No.:	382
	SDG No.:	LK7349

The attached data report contains the analytical results of samples that were submitted to Lockheed Analytical Services on 29 June 1996.

The temperature of the cooler upon receipt was 4°C. Sample containers received agree with the chain-of-custody documentation. Sample containers were received intact. Samples designated for hexachrome analysis were not received in time to meet the analytical holding time requirements.

The case narratives included in the following attachments provide a detailed description of all events that occurred during sample preparation, analysis, and data review specific to the samples and analytical methods requested.

A list of data qualifiers, chain-of-custody forms, sample receiving checklist, and log-in report are also enclosed representing the samples received within this group.

If you have any questions concerning the analysis or the data please call Kathleen Hall at (509) 375-4741.

Lockheed Analytical Services

Log-in No.: L7349
Quotation No.: Q400000-B
SAF: B96-092
Document File No.: 0629596
BHI Document File No.:382
SDG No.: LK7349
Page1

Release of this data report has been authorized by the Laboratory Director or the Director's designee as evidenced by the following signature.

" I certify that this data package is in compliance with the SOW, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Manger or a designee, as verified by the following signature."

Sincerely,

Karen Hermann for

Kathleen M. Hall
Client Services Representative

cc: Client Services
Document Control

CASE NARRATIVE
INORGANIC NON METALS ANALYSES

The routine calibration and quality control analyses performed for this batch include as applicable: initial and continuing calibration verification, initial and continuing calibration blanks, method blank(s), laboratory control sample(s), matrix spike (predigestion) sample(s), duplicate sample(s).

Preparation and Analysis Requirements

- One water sample was received for LK7349 and analyzed in batch 629bh for selected analytes as requested on the chain of custody. Quality control analysis was performed on the following samples:

Client ID	LAL #		Method
BOHD40	L7349-2	DUP,MS	7196 Chromium (VI)

Holding Time Requirements

- All samples were received and analyzed outside of the method-specific holding times and the associated sample is flagged with an "H".

Internal Quality Control

- All Internal Quality Control were within acceptance limits.

Kay McCann
Prepared By

July 9, 1996
Date

CASE NARRATIVE
INORGANIC METALS ANALYSES

The routine calibration and quality control analyses performed for this batch include as applicable: instrument tune (ICP/MS only), initial and continuing calibration verification, initial and continuing calibration blanks, method blank(s), laboratory control sample(s), ICP interference check samples (ICP only), serial dilutions, analytical (post-digestion) spike samples, matrix spike (predigestion) sample(s), duplicate sample(s).

Preparation and Analysis Requirements

All samples were received on June 29, 1996. The samples were logged in as L7349 and were prepared and analyzed in batch 629 bh. The samples were analyzed by Method 200.7 ICP Metals.

Holding Time Requirements

- All samples were analyzed within the method-specific holding times.

Method Blanks

- The concentration levels of all the requested analytes in the method blank were below the reporting detection limits.

Internal Quality Control

- All Internal Quality Control were within acceptance limits.

Shellee McGrath
Prepared By

July 24, 1996
Date

LOCKHEED ANALYTICAL SERVICES
LOGIN CHAIN OF CUSTODY REPORT (ln01)
Jun 29 1996, 12:01 pm

Login Number: L7349
Account: 596 Bechtel Hanford, Inc. * Richland, WA
Project: BECHTEL-HANFORD Bechtel Hanford Project

Laboratory Sample Number	Client Sample Number	Collect Date	Receive Date	Due PR Date
L7349-1	B0HD40	27-JUN-96	29-JUN-96	03-AUG-96
TEMP 4				
Location: 157				
Water 1 S SCREENING		Hold:24-DEC-96		
L7349-2	B0HD40	27-JUN-96	29-JUN-96	03-AUG-96
TEMP 4				
Location: RFG02-25A				
Water 1 S 7196 CHROMIUM (VI)		Hold:28-JUN-96		
L7349-3	B0HD40	27-JUN-96	29-JUN-96	03-AUG-96
TEMP 4				
Location: RFG02-25A				
Water 1 S 200.7 METALS		Hold:24-DEC-96		
L7349-4	REPORT TYPE	29-JUN-96	29-JUN-96	03-AUG-96
Location:				
Water 1 S EDD - DISK DEL.				
Water 1 S INORG TYPE 4A RPT				

Signature: Paul Dans

Date: 6-29-96 0009

01+29596

Bechtel Hanford, Inc.	CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST L7349	Page 1 of 1 Data Turnaround ☐ Priority ■ Normal

Collector	Company Contact	Telephone
R.Fahlberg	M.T. Stankovich	372-9626
Project Designation	**Sampling Location**	**SAF No.**
100-HR-3 Routine Process Samples	100 Area	B96-092
Ice Chest No.	**Field Logbook No.**	**Method of Shipment**
LRD-C10	EL-1309	Hqnd Delivered
Shipped To	**Offsite Property No.**	**Bill of Lading/Air Bill No.**
Lockheed	~~NA-BW~~ 6/96-0-0314-1	~~NA-BN~~ 2904659265

Possible Sample Hazards/Remarks	Preservation	HNO3	cool to 4c	None								
	Type of Container	G/P	G/P	G/P								
	No. of Containers	1	1	1								
Special Handling and/or Storage	Volume	500mL	500mL	20mL								
SAMPLE ANALYSIS		ICP Metals, 2 Cr	Cr Hex	Activity Scan								

Sample No.	Matrix*	Date Sampled	Time Sampled										
B0HD40	W	6/27/96	0935	X	X	X			Eff	SP300			

CHAIN OF POSSESSION	Sign/Print Names	
Relinquished By	Date/Time 1210 6-27-96	Received By ER Date/Time 1210 6-27-96
Relinquished By ER	Date/Time 0800 6-28-96	Received By Date/Time
Relinquished By	Date/Time	Received By Date/Time
Relinquished By	Date/Time	Received By Date/Time

SPECIAL INSTRUCTIONS
Sample analysis for Chromium VI is requested for information only. The ERC contractor acknowledges the 24-hour holding time will not be met.

Matrix*
S = Soil
SE = Sediment
SO = Solid
SL = Sludge
W = Water
O = Oil
A = Air
DS = Drum Solids
DL = Drum Liquids
T = Tissue
WI = Wipe
L = Liquid
V = Vegetation
X = Other

LABORATORY SECTION	Received By	Title	Date/Time
FINAL SAMPLE DISPOSITION	Disposal Method	Disposed By	Date/Time

Environmental Restoration Contractor **ERC Team**

Interoffice Memorandum

Job No. 22192
Written Response Required: NO
CCN: N/A
OU: N/A
TSD: N/A
ERA: N/A
Subject Code: 5830

TO:	W. S. Thompson	N1-28	DATE:	February 29, 1996
	G. C. Henckel	H4-80		
COPIES:	K. A. Smith	X0-23	FROM:	S. K. De Mers
	T. L. Lafreniere	X0-23		Radiological Controls
	D. E. Gergely	X0-23		T7-05/373-1913

SUBJECT: **Total Activities for Off-Site Shipments of Groundwater Samples to NRC Licensed Laboratories**

There is no need to perform total activities prior to offsite shipment to NRC licensed labs of samples taken from ground water wells located on the Hanford Site.

All wells reviewed to date for radiological content have shown no well with a total activity in excess of 2,000,000 pCi/l (2,000 pCi/gm), the Department Of Transportation limit for radioactive material. The highest activity in any known well is 1.56×10^6 pCi/l H^3.

While this does not constitute any release from radiological controls for worker protection, it does allow samples to be shipped based on historical laboratory data and save the expense of doing radiochemical analysis.

A copy of the most recent analytical data should be provided to the NRC licensed laboratory with the samples being shipped or if no data is available for new wells, the most recent data from adjacent wells.

LOCKHEED MARTIN

Sample Login
Login Review Checklist

Lot Number L7349

The login review should be conducted by that person logging in the samples as well as a peer. Please use this checklist to ensure that such reviews occur in a uniform basis. Please sign and date below to verify that a login review has occurred. This checklist should be affixed to each login package prior to distribution.

For effective login review, at a minimum, five reports form the login process are required. These are the COC (or equivalent), the login COC report, the sample summary report, the sample receiving checklist, and the login quotation. Before beginning review, ensure that these five components are available. Jobs with single component samples, the sample summary report may be omitted.

SAMPLE SUMMARY REPORT	YES	NO	N/A	Comment
1. Are all sample ID's correct?	X			
2. Are all samples present?	X			
3. Are all matrices indicated correctly?	X			
4. Are all analyses on the COC logged in for the appropriate samples?	X			
5. Are all analyses logged in for the correct container?	X			
6. Are samples logged in according to LAS batching procedures?	X			

LOGIN CHAIN OF CUSTODY	YES	NO	N/A	Comment
1. Are the collect, receive, and due dates correct for every sample?	X			
2. Have all appropriate comments been indicated in the comment section?	X			

SAMPLE RECEIVING CHECKLIST	YES	NO	N/A	Comment
1. Are all discrepancies between the COC and the login noted (if applicable)?	X			

[signature] 6-24-96
primary review signature date

[signature] 6-24-96
secondary review signature date

0012006459C

SAMPLE CHECK-IN LIST

Date/Time Received: 6-29-96 / 9:45 SDG#: 11.2

Work Order Number: 11.4 SAF #: 97k-092

Shipping Container ID: 1100-516 Chain of Custody # 11.3

1. Custody Seals on shipping container intact? Yes [x] No [ ]
2. Custody Seals dated and signed? Yes [x] No [ ]
3. Sample temperature 4i
4. Vermiculite/packing materials is Wet [ ] Dry [x]
5. Each sample is in a plastic bag? Yes [x] No [ ]
6. Sample holding times exceeded? Yes [x] No [ ]

7. Samples have:
   ____tape ____hazard labels
   x custody seals ____appropriate sample labels

8. Samples are:
   x in good condition ____leaking
   ____broken ____have air bubbles

9. Is the information on the COC and Sample bottles in agreement?

   Yes[x] No [ ]

Notes: This sample was passed its 24hr Holding Time

Sample Custodian/Laboratory: [illegible] Date: 6-29-96

~~Telephoned~~ (faxed) To: Kathleen Hall On: 6-29-96 By Paul C Davis

Was 6-29 °C

0013

[illegible]

Lockheed Analytical Services
Sample Receiving Checklist

Client Name: [illegible] **Job No.** L7349 **Cooler ID:** [illegible]

COOLER CONDITION UPON RECEIPT			
Temperature of cooler upon receipt:	4°C		
temperature of temp. blank upon receipt:	—		
	Yes	No	* Comments/Discrepancies
custody seals intact	X		
chain of custody present	X		
blue ice (or equiv.) present/frozen	X		
rad survey completed	X		

SAMPLE CONDITION UPON RECEIPT			
	Yes	No	* Comments/Discrepancies
all bottles labeled	X		
samples intact	X		
proper container used for sample type	X		
sample volume sufficient for analysis	X		
proper pres. indicated on the COC	X		
VOA's contain headspace		N/A	
are samples bi-phasic (if so, indicate sample ID'S):		N/A	

MISCELLANEOUS ITEMS			
	Yes	No	* Comments/Discrepancies
samples with short holding times	X		Chromium Hex passed the 24hr
samples to subcontract			hold time client acknowledges this discrepancy

ADDITIONAL COMMENTS/DISCREPANCIES

Completed by / date:

Sent to the client (date/initials): ** Client's signature upon receipt:

*Notes: * = contact the appropriate CSR of any discrepancies immediately upon receipt*

*** = please review this information and return via facsimille to the appropriate CSR (702) 361-3146*

001

version 2.0 (11/11/94)

Lockheed Analytical Laboratory
SAMPLE SUMMARY REPORT (su02)
Bechtel Hanford, Inc. * Richland, WA

Client Sample Number	LAL Sample Number	SDG Number	Matrix	Method
B0HD40	L7349-1		Water	SCREENING
	L7349-2		Water	7196 CHROMIUM (\
	L7349-3		Water	200.7 METALS
REPORT TYPE	L7349-4		Water	EDD - DISK DEL.
	L7349-4		Water	INORG TYPE 4A RI

LOCKHEED ANALYTICAL SERVICES

Sample Results

Client Sample ID: B0HD40	Date Collected: 27-JUN-96
Matrix: Water	Date Received: 29-JUN-96
Percent Solids: N/A	

Constituent	Units	Method	Result	Project Reporting Limit	Data Qualifier(s)	Date Analyzed	LAS Batch ID	LAS Sample ID
Chromium, hexavalent	mg/L	7196	< 0.003	0.020	HU	08-JUL-96	38718	L7349-2

Lockheed Analytical Laboratory

Determination of Hexavalent Chromium Calibration and Calibration Verification Results

LAL Batch ID: 629-BH
Work Group: 7196 CHROMIUM (VI)_38718
Method: 7196 (Hexavalent Chromium)

Calibration Results

Standard Concentration (mg/L)	Measured Instrument Response	Linearized Instrument Response	Calculated Concentration (mg/L)	Standard Recovery (%)
0.000	0.000	0.000	-0.002	
0.025	0.021	0.021	0.025	99
0.050	0.043	0.043	0.052	105
0.100	0.081	0.081	0.100	100
0.200	0.161	0.161	0.200	100
0.250	0.200	0.200	0.249	100

Slope = 1.2544
Intercept = -0.0016
Correlation (r) = 0.9999

Measured Instrument Response: Absorbance (540 nm)

Calibration Verification Results

Sample Identification	True Concentration (mg/L)	Found Concentration (mg/L)	Analyte Recovery (%)
ICV	0.1	0.101	101
CCV	0.1	0.100	100

Calibration Blank Results

Sample Identification	Analyte Found (mg/L)
ICB	0.003 U
CCB	0.003 U

Lockheed Analytical Laboratory

Determination of Hexavalent Chromium
Quality Control Results

LAL Batch ID: 629-BH
Work Group: 7196 CHROMIUM (VI)_38718
Method: 7196 (Hexavalent Chromium)

Laboratory Control Sample/Duplicate Results (Recovery)

Sample Identification	True Concentration (mg/L)	Found Concentration (mg/L)	Analyte Recovery (%)
LCS	0.05	0.055	110
LCSD		(No LCSD analyzed)	

Laboratory Control Sample/Duplicate Results (Difference)

LCS Result (mg/L)	LCSD Result (mg/L)	Relative Difference (%)	Flag
	(No LCSD analyzed)		

Preparation Blank Results

Sample Identification	Analyte Found (mg/L)
PB	0.003 U

Sample Duplicate Results (Difference)

LAL Sample Identification	Sample Result (mg/L)	Duplicate Result (mg/L)	Relative Difference (%)	Flag
L7349-2	0.003 U	0.003 U		b

Spiked Sample/Spike Duplicate Results (Recovery)

LAL Sample Identification	Sample Result (mg/L)	Analyte Added (mg/L)	Spike Result (mg/L)	Spike Recovery (%)	Flag
L7349-2S	0.003 U	0.05	0.056	112	

Spiked Sample/Spike Duplicate Results (Difference)

Spike Result (mg/L)	Spike Dup Result (mg/L)	Relative Difference (%)	Flag
	(No spike duplicate analyzed)		

LOCKHEED ANALYTICAL SERVICES

Sample Results

Client Sample ID: B0HD40	Date Collected: 27-JUN-96
Matrix: Water	Date Received: 29-JUN-96
Percent Solids: N/A	

Constituent	Units	Method	Result	MDL	RDL	Data Qual	Dilution	Date Analyzed	LAS Batch ID	LAS Sample ID
CHROMIUM	mg/L	200.7	0.0077	0.0060	0.010	B	1	16-JUL-96	38760	L7349-3